

**Izvešće o ispitivanju br.  
Test Report No.**

**16 TR 0040**

Predmet ispitivanja: **Enclosure with probe**  
Subject of testing:

Tipna oznaka: **Integral Flo-Tel xd**  
Type identification:

Proizvođač: **ELFAB LTD., Alder road, North shields, Tyne & Wear, NE29 8SD,**  
Manufacturer: **England**

Podnositelj zahtjeva: **ELFAB LTD., Alder road, North shields, Tyne & Wear, NE29 8SD,**  
Applicant: **England**

**Rezultati ispitivanja odnose se samo na ispitani uzorak.  
The test results reported are relevant only for the tested sample(s).**

Ovo izvješće ima 3 stranica i - stranica priloga.  
This report comprises 3 pages and - pages of appendices.

**Ovo izvješće o ispitivanju može se umnažati samo u cijelosti i bez izmjena,  
uključujući i priloge.**

**This test report may only be reproduced in its entirety and without any change,  
appendices included.**

Broj posla / Job Id. No.: **PB.16.TI.353/SS**

Datum izrade: **26.04.2016**  
Date of issue:

Voditelj laboratorija:  
Head of laboratory:



**M. Mačković, mag.ing.univ.spec.mech.**



**1 Standards applied:**

HRN EN 60529-0:2000/A1:2008 EN 60529:1991/A1:2000 IEC 60529:1989/am1:1999 Ed. 2.0

**2 Description and identification of sample(s) submitted for test:**

Sample identification	No. of pcs	Sample description	Remark
Integral Flo-Tel xd	1	Enclosure with probe prepared for testing, drawing no. 05/01/08773 SHT1	-

Note 1: Samples are recorded in the main samples receipt book PQI-C-106, book No. (year): 2016, as follows: 33 dated 20.04.2016.




Photo no.1: Tested sample Integral Flo-Tel xd

**Order of tests was as follows:**

Test No.	Standard reference	Clause	Title	Sample(s) tested
1	IEC 60529	13.4	<i>Tests for protection against solid foreign objects indicated by the first characteristic numeral 6</i>	Integral Flo-Tel xd
2	IEC 60529	14.2.6	<i>Tests for protection against water indicated by the second characteristic numeral 6</i>	Integral Flo-Tel xd



